

Authorized Service Agent -
CGG Service do Brasil LTDA
 Niterói / RJ - BRAZIL
 Tel: +55 21 3254-8895
 e-mail: contato@cggservice.com.br
 website: www.cggservice.com.br

On behalf of -
RWO Veolia Water Technology GmbH
 Gerold-Janssen-Straße 2
 Bremen - Germany
 Tel: +49 421 53705-0
 e-mail: service.rwo@veoliawater.com
 website: www.rwo.de
IMO MARPOL MEPC.107(49) APPROVED

Calibration Check Certificate for 15PPM Bilge Alarm

Certificate No.: CE221124A

Customer:	BRAM OFFSHORE		
Ship Name/IMO:	C VIKING	/	9640231
Measuring Cell:	RWO OMD-24	Serial No.: 4015244	Manuf.: MAR.2018
OMD Computer:	RWO OMD-24	Serial No.: 2008439	Manuf.: MAY/2016
OWS Model:	RWO SKIT/S-DEB 1.0m³/h	Serial No.: 12814	Manuf.: 2011
Voltage:	Equip: 440V60HZ	Meas. Cell: 24V 1AC	
Instruments:	Multimeter Hikary Class IV Isolation (INMETRO Calibrated). Test Fluid Type A, Lot No. JL0076.2211, complies with IMO MEPC.107(49) Resolutions.		

Tests Results:

	Value	First Check	Second Check	Status
Clean Water Temperature	15 - 40 °C	33 °C	33 °C	OK
Check With Distilled Water	0 (+/-)PPM	0 PPM	0 PPM	OK
Check With Test Fluid Concentration in Distilled H ₂ O	Value	First Check	Second Check	Status
15 drops	15 (+/-5)PPM	16 PPM	16 PPM	OK

Remarks: The Instrument has been tested according to the instructions of the manufacturer and complies with the requirements specified in IMO Resolutions MEPC.107(49).

Validation: This Certificate is valid until **24.NOVEMBER.2023**

Agent Confirms:



Eng. Gerson GARCIA
 MANAGING DIRECTOR
 CGG Service do Brasil LTDA.

Customer's Confirms (Stamp and Signature):

RENATO PEDROSO DE ALBUQUERQUE
 CHEFE DE MÁQUINA-10M
 CIR 381P2001296466

CGG Service do Brasil

Technician: Gerson GARCIA

24.NOV.2022
 Date